

# CalCOFI 0711NH: CTD Oxygen vs Bottle Oxygen

4sec CTD sensor O2 vs bottle O2 (ml/L)

$$y = 0.9865x - 0.0057$$
$$R^2 = 0.9998$$

