

CalCOFI 0907M2: Primary CTD Oxygen vs Bottle Oxygen

4sec average primary sensor O2 vs bottle O2 (ml/L)

$$y = 1.0622x - 0.086$$
$$R^2 = 0.9994$$

