

CalCOFI 0907M2: Secondary CTD Oxygen vs Bottle Oxygen

4sec avg secondary sensor O2 vs bottle O2 (ml/L)

$$y = 1.0375x - 0.0345$$
$$R^2 = 0.9997$$

