

CalCOFI 0907M2: Secondary CTD Oxygen vs Bottle Oxygen

4sec avg secondary sensor O2 vs bottle O2 (ml/L)

