

CalCOFI 1110NH: CTD Primary O2 vs Bottle O2

4-sec ave primary CTD O2 sensor vs bottle oxygen (umol/Kg)

$$y = 1.0237x - 0.6739$$
$$R^2 = 0.9993$$

