

CalCOFI 1110NH: CTD Secondary O2 vs Bottle O2

4-sec ave secondary CTD O2 sensor vs bottle oxygen (ml/L)

$$y = 1.0438x - 0.0113$$
$$R^2 = 0.9993$$

