

CalCOFI 1110NH: CTD Secondary O2 vs Bottle O2

4-sec ave secondary CTD O2 sensor vs bottle oxygen (umol/Kg)

$$y = 1.0459x - 0.3808$$
$$R^2 = 0.9993$$

