

CalCOFI 1501NH: Primary CTD Oxygen vs Bottle Oxygen

4sec ave Ox1 sensor oxygen vs final bottle O2 (uM/Kg)

$$y = 1.0507x - 0.4099$$
$$R^2 = 0.9997$$

