

CalCOFI 1501NH: Secondary CTD Oxygen vs Bottle Oxygen

4sec ave Ox2 sensor oxygen vs final bottle O2 (uM/Kg)

$$y = 1.0506x - 0.5119$$
$$R^2 = 0.9997$$

