

# CalCOFI 1507OC: Secondary CTD Oxygen vs Bottle Oxygen

4 sec ave CTD sensor vs bottle O<sub>2</sub> (umol/Kg)

$$y = 1.0399x - 0.5543$$

$$R^2 = 0.9996$$

