

# CalCOFI 1601RL: Primary CTD Oxygen vs Bottle Oxygen

4-sec ave sensor oxygen vs bottle sample (ml/L)

$y = 1.0362x - 0.0032$   
 $R^2 = 0.9998$

