

CalCOFI 1708SR: Secondary CTD Oxygen vs Bottle Oxygen

4sec ave secondary CTD O2 sensor vs bottle O2 (ml/L)

$$y = 1.0421x + 0.0058$$
$$R^2 = 0.9996$$

