

CalCOFI 1708SR: Secondary CTD Oxygen vs Bottle Oxygen

4sec ave secondary CTD O2 sensor vs bottle O2 (uM/Kg)

$$y = 1.0429x + 0.3795$$
$$R^2 = 0.9996$$

