

CalCOFI 1802SH: Primary CTD Oxygen vs Bottle Oxygen

4sec avg primary CTD sensor O2 vs bottle O2 (ml/L)

$$y = 1.0727x + 0.0021$$
$$R^2 = 0.9997$$

