

# SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0357  
CALIBRATION DATE: 20-May-88

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## GHIJ COEFFICIENTS

g = -4.32706875e+000  
h = 5.05711491e-001  
i = -1.51690511e-004  
j = 3.35943316e-005  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 4.99350461e-007  
b = 5.06154591e-001  
c = -4.33235095e+000  
d = 9.10855328e-004  
m = 5.4  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.92558	0.00000	0.00000
-1.0058	35.0475	2.82096	8.01370	2.82100	0.00004
7.0030	35.0482	3.53310	8.84431	3.53304	-0.00006
14.9753	35.0896	4.29914	9.65735	4.29913	-0.00001
23.0270	35.0891	5.11362	10.45142	5.11367	0.00005
31.0019	35.0873	5.95618	11.21245	5.95616	-0.00002

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10 (1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

