

# SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0722  
CALIBRATION DATE: 06-Aug-98

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## GHIJ COEFFICIENTS

g = -3.84388655e+000  
h = 4.15117521e-001  
i = 8.16186868e-005  
j = 1.68850176e-005  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 3.63072951e-005  
b = 4.15334986e-001  
c = -3.84475162e+000  
d = -8.67988098e-005  
m = 3.8  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	3.04150	0.00000	0.00000
-1.4107	35.0095	2.78379	8.71522	2.78380	0.00000
1.1279	35.0106	3.00213	9.00935	3.00213	0.00001
15.2473	35.0102	4.31722	10.60730	4.31720	-0.00002
18.6852	35.0083	4.65974	10.98474	4.65973	-0.00001
29.2278	35.0053	5.75404	12.11044	5.75411	0.00006
32.6673	35.0012	6.12265	12.46608	6.12261	-0.00004

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

