

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0722
CALIBRATION DATE: 11-Mar-03

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -3.84545560e+000
h = 4.15583390e-001
i = 1.48744438e-005
j = 1.94273424e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 2.63814948e-005
b = 4.15571864e-001
c = -3.84521507e+000
d = -8.02499740e-005
m = 3.9
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	3.04108	0.00000	0.00000
-0.5002	32.7160	2.69078	8.58622	2.69077	-0.00000
0.9998	32.7161	2.81287	8.75441	2.81287	0.00000
14.9998	32.7174	4.04031	10.29147	4.04030	-0.00001
18.4998	32.7172	4.36878	10.66470	4.36880	0.00002
28.9998	32.7158	5.39550	11.75347	5.39547	-0.00003
32.4998	32.7102	5.74864	12.10487	5.74865	0.00002

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

