

# SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0722  
CALIBRATION DATE: 12-Aug-04

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## GHIJ COEFFICIENTS

g = -3.84502136e+000  
h = 4.15555114e-001  
i = 1.78043248e-005  
j = 1.99525164e-005  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 2.73137646e-005  
b = 4.15552451e-001  
c = -3.84479489e+000  
d = -8.05014471e-005  
m = 3.9  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	3.04096	0.00000	0.00000
-1.0003	34.8612	2.80783	8.74706	2.80783	0.00000
0.9997	34.8615	2.97943	8.97818	2.97944	0.00000
14.9997	34.8628	4.27668	10.56052	4.27667	-0.00002
18.4997	34.8628	4.62384	10.94457	4.62385	0.00001
28.9997	34.8618	5.70890	12.06456	5.70891	0.00001
32.4997	34.8573	6.08230	12.42605	6.08229	-0.00001

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

