

# SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2206  
CALIBRATION DATE: 02-Mar-01

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## GHIJ COEFFICIENTS

g = -1.05214061e+001  
h = 1.66682853e+000  
i = -2.10556988e-003  
j = 2.48671111e-004  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 5.67464668e-007  
b = 1.66156887e+000  
c = -1.05118774e+001  
d = -8.57335229e-005  
m = 6.3  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.51523	0.00000	0.00000
-1.4396	35.0010	2.78073	4.80160	2.78071	-0.00002
1.0209	35.0010	2.99206	4.93217	2.99208	0.00002
15.1759	35.0016	4.30924	5.67825	4.30922	-0.00002
18.6521	35.0008	4.65551	5.85856	4.65554	0.00003
29.0606	34.9992	5.73534	6.38787	5.73532	-0.00003
32.6182	34.9925	6.11599	6.56421	6.11601	0.00002

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

