

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2206
CALIBRATION DATE: 11-Mar-05

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -1.05292030e+001
h = 1.67003338e+000
i = -3.05004706e-003
j = 3.19244094e-004
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 3.57931437e+001
b = -4.27680117e+001
c = 3.76333651e+001
d = 4.96753031e-001
m = 2.1
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.51519	0.00000	0.00000
-1.0005	34.9440	2.81386	4.82237	2.81386	-0.00000
0.9995	34.9447	2.98585	4.92845	2.98585	-0.00000
14.9995	34.9456	4.28574	5.66603	4.28575	0.00001
18.4995	34.9459	4.63365	5.84756	4.63364	-0.00001
28.9995	34.9442	5.72085	6.38134	5.72085	0.00000

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10 (1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

