




CERTIFICATE OF CALIBRATION AND TESTING

CUSTOMER: NOAA
P.O. NUMBER: (G Streeter)
MODEL: 85000
SERIAL NUMBER: UB00981



Speed (m/s)	Avg Diff	Std Dev
5.98	0.1%	1.1%
19.04	0.0%	0.3%
31.58	-0.3%	0.3%

R. M. Young Company certifies that the above equipment was inspected and calibrated prior to shipment in accordance with established manufacturing and testing procedures. Standards established by R.M. Young Company for calibrating the measuring and test equipment used in controlling product quality are traceable to the National Institute of Standards and Technology.

To maintain these specifications, regular maintenance intervals are essential.

Calibration Date: 6-Dec-2012

R.M. Young Company

Ryan Phillips

R.M. YOUNG COMPANY 2801 Aero Park Drive, Traverse City Michigan 49686-9171 USA
Tel: 231-946-3980 Fax: 231-946-4772 Email: met.sales@youngusa.com
ISO 9001:2008 CERTIFIED