

Biospherical Instruments Inc

CALIBRATION CERTIFICATE

UNDERWATER PAR SENSOR WITH LOG AMPLIFIER

Calibration Date: 07/08/10

Job No.: R-10648

Model Number: QSP-200L

Serial Number: 4329

Operator: TPC

Standard Lamp: GS1024(8/28/08)

Operating Voltage Range: 6 to 15 VDC (+)

Note: The QSP-200 uses a log amplifier to measure the detector signal current with $V = \log I (\text{Amps}) / I_{\text{Ref}}$
To calculate irradiance, use this formula:

Irradiance = Calibration factor * (10^{Light Signal Voltage} - 10^{Dark Voltage})
--

With the appropriate (solar corrected) Irradiance Calibration Factor:-

Dry Calibration Factor:	1.52E+13	quanta/cm²·sec/"amps"	2.53E-05	μEinsteins/cm²·sec/"amps"
Wet Calibration Factor:	2.56E+13	quanta/cm²·sec/"amps"	4.25E-05	μEinsteins/cm²·sec/"amps"

Sensor Test Data and Results⁴⁾

Sensor Supply Current (Dark):		81.1	mA							
Supply Voltage:		6	Volts							
Lamp Integrated PAR Irradiance:		9.27E+15	quanta/cm ² ·sec	0.01540	μEinsteins/cm ² ·sec					
SC3 Immersion Coefficient:		0.594	Scalar Correction:	1	PAR Solar Correction:		1.0000			
Nominal Filter OD	Calibrated Trans.	Sensor Voltage	Measured Trans.	Measured Signal (Amps)	Estimated Signal (Amps)	Calc. Output (Volts)	Error (Volts)	Error (%)	Test Irrad. (quanta/cm ² ·sec)	
No Filter	100.00%	2.786	100.00%	6.11E-08	6.11E-08	2.787	0.001	0.0	9.27E+15	
0.3	36.10%	2.345	36.09%	2.21E-08	2.21E-08	2.346	0.001	0.0	3.35E+15	
0.5	27.60%	2.231	27.71%	1.69E-08	1.69E-08	2.230	-0.001	-0.4	2.57E+15	
1	9.27%	1.766	9.35%	5.71E-09	5.66E-09	1.763	-0.003	-0.9	8.67E+14	
2	1.11%	0.913	1.13%	6.93E-10	6.78E-10	0.906	-0.007	-2.2	1.05E+14	
3	0.05%	0.238	0.08%	4.68E-11	3.26E-11	0.201	-0.037	-30.2	7.09E+12	

Dark Before:	<u>0.100</u>	Volts		
Light - No Filter Hldr.:	<u>2.785</u>	Volts	$I_{\text{Ref}} =$	<u>1.00E-10</u> Amps
Dark After - NFH:	<u>0.103</u>	Volts	$I_{\text{Dark}} =$	<u>1.26E-10</u> Amps
Average Dark	<u>0.102</u>	Volts	$10^{V_{\text{Dark}}} =$	<u>1.263281</u> Amps

RG780 **0.139**

Notes:

1. Annual calibration is recommended.
2. There is increasing error associated with readings below zero.
3. The collector should be cleaned frequently with alcohol.
- 4) This section is for internal use and for more advanced analysis.