



# THE EPPLEY LABORATORY, INC.

12 Sheffield Avenue, PO Box 419, Newport, Rhode Island USA 02840  
Phone: 401.847.1020 Fax: 401.847.1031 Email: info@eppleylab.com

## Calibration Certificate

---

Instrument: Precision Spectral Pyranometer, Model PSP, Serial Number 29098F3

Procedure: This pyranometer was compared in Eppley's Integrating Hemisphere according to procedures described in *ISO 9847 Section 5.3.1* and Technical Procedure, TP01 of The Eppley Laboratory, Inc.'s Quality Assurance Manual on Calibrations.

Transfer Standard: Eppley Precision Spectral Pyranometer, Model PSP, Serial Number 18851F3

Results: **Sensitivity:**  $S = 7.46 \mu\text{V} / \text{Wm}^{-2}$   
Uncertainty:  $U_{95} = \pm 0.91\%$  (95% confidence level,  $k=2$ )  
Resistance:  $682 \Omega$  at  $23^\circ\text{C}$

Date of Test: February 14, 2014

Traceability: This calibration is traceable to the World Radiation Reference (WRR) through comparisons with Eppley's AHF standard self-calibrating cavity pyrheliometers which participated in the Eleventh International Pyrheliometric Comparisons (IPC XI) at Davos, Switzerland in September-October 2010. Unless otherwise stated in the remarks section below or on the Sales Order, the results of this calibration are "AS FOUND / AS LEFT".

Due Date: Eppley recommends a minimum calibration cycle of five (5) years but encourages annual calibrations for highest measurement accuracy.

Customer: Oregon State University  
Corvallis, OR

Signatures: *Debra L. Bivins* In Charge of Test: *Thomas D. Kerk* Reviewed by:

Eppley SO 64023

Date of Certificate February 17, 2014

Remarks: Sensitivity before Repainting Element =  $8.05 \mu\text{V} / \text{Wm}^{-2}$