PO Box 518 620 Applegate St. Philomath, OR 97370



(541) 929-5650 Fax (541) 929-5277 www.wetlabs.com

WETStar Calibration and Repairs

Date January 21, 2009 Customer

Oregon State

S/N#

WS3S-942P

Repair Order 5531

Standard Service

· Performed noise test: 1 sample/sec for 60 sec

• Performed stability test: 1 sample/min for 12 hrs

Performed temperature test: 25–2 °C

- · Performed saturation test
- · Shake-tested unit
- · Pressure-tested unit
- · Updated unit's calibration sheet

Additional Repairs

Replaced Bulkhead Connector and O-Rings.

Comments

4/17/08 Revision I wsxwkbka1.xls