

M1000 Modules H, I, J, K, L, and M Calibration Check

	Input mV	module H	H %error	module I	I %error	module J	J %error	module K	K %error	module L	L %error
Record 0	0.00	0.19	-0.1920	0.00	0.0000	-0.15	0.1493	0.27	-0.2667	0.25	-0.2507
Record 1	0.00	0.16	-0.1600	0.00	0.0000	-0.09	0.0853	0.25	-0.2453	0.32	-0.3200
Record 2	11.11	11.38	-2.3942	11.28	-1.5302	11.27	-1.4341	11.41	-2.6823	11.34	-2.0582
Record 3	111.11	111.18	-0.0594	110.97	0.1278	110.91	0.1830	111.13	-0.0162	111.14	-0.0234
Record 4	222.22	222.31	-0.0402	221.98	0.1086	222.16	0.0270	222.18	0.0174	222.25	-0.0114
Record 5	333.33	333.35	-0.0074	333.13	0.0598	333.28	0.0150	333.59	-0.0778	333.37	-0.0106
Record 6	444.44	444.53	-0.0210	444.32	0.0270	444.21	0.0510	444.39	0.0102	444.01	0.0966
Record 7	555.55	555.36	0.0342	555.10	0.0803	555.58	-0.0061	555.62	-0.0119	555.46	0.0169
Record 8	666.66	666.68	-0.0026	666.56	0.0150	666.62	0.0054	666.46	0.0294	666.67	-0.0010
Record 9	777.77	777.76	0.0013	777.78	-0.0015	777.80	-0.0042	778.13	-0.0460	778.09	-0.0412
Record 10	888.88	888.61	0.0306	888.85	0.0030	888.68	0.0222	888.96	-0.0090	889.12	-0.0270
Record 11	999.99	999.54	0.0449	999.70	0.0289	999.68	0.0310	999.83	0.0155	1000.43	-0.0437
Record 12	1111.11	1110.80	0.0284	1110.83	0.0255	1110.69	0.0380	1111.09	0.0015	1111.33	-0.0196
Record 13	2222.22	2221.76	0.0207	2222.22	0.0000	2221.44	0.0351	2222.20	0.0010	2222.43	-0.0095
Record 14	3333.33	3332.53	0.0239	3332.94	0.0117	3332.12	0.0364	3333.04	0.0089	3333.93	-0.0180
Record 15	4444.44	4443.48	0.0217	4444.72	-0.0062	4442.92	0.0341	4443.40	0.0233	4444.61	-0.0038
Record 16	4955.55	4954.50	0.0213	4955.41	0.0028	4953.78	0.0357	4954.57	0.0198	4955.61	-0.0012
Record 17	-4955.55	-4955.37	0.0037	-4955.41	0.0028	-4954.92	0.0127	-4955.29	0.0053	-4956.28	-0.0148
Record 18	-4444.44	-4444.05	0.0087	-4444.29	0.0034	-4443.76	0.0154	-4444.02	0.0095	-4444.97	-0.0118
Record 19	-3333.33	-3332.90	0.0129	-3333.64	-0.0092	-3332.97	0.0110	-3333.10	0.0069	-3333.81	-0.0145
Record 20	-2222.22	-2221.85	0.0169	-2222.34	-0.0054	-2221.95	0.0121	-2222.11	0.0051	-2222.39	-0.0076
Record 21	-1111.11	-1111.09	0.0020	-1111.36	-0.0221	-1110.88	0.0207	-1111.13	-0.0019	-1110.96	0.0140
Record 22	-999.99	-999.92	0.0070	-1000.23	-0.0239	-999.67	0.0315	-999.86	0.0134	-999.86	0.0129
Record 23	-888.88	-888.65	0.0264	-889.21	-0.0366	-888.50	0.0426	-888.71	0.0186	-888.85	0.0030
Record 24	-777.77	-777.62	0.0198	-778.04	-0.0344	-777.83	-0.0083	-777.63	0.0177	-777.85	-0.0097
Record 25	-666.66	-666.58	0.0118	-666.62	0.0054	-666.41	0.0382	-666.69	-0.0050	-666.65	0.0014
Record 26	-555.55	-555.58	-0.0047	-555.67	-0.0215	-555.56	-0.0013	-555.64	-0.0167	-555.70	-0.0277
Record 27	-444.44	-444.53	-0.0210	-444.33	0.0246	-444.47	-0.0066	-444.64	-0.0450	-444.64	-0.0450
Record 28	-333.33	-333.35	-0.0058	-333.52	-0.0570	-333.37	-0.0114	-333.40	-0.0210	-333.52	-0.0570
Record 29	-222.22	-222.03	0.0846	-222.06	0.0702	-222.31	-0.0402	-222.21	0.0054	-222.32	-0.0450
Record 30	-111.11	-111.17	-0.0522	-111.20	-0.0810	-111.19	-0.0690	-111.04	0.0630	-111.13	-0.0162
Record 31	-11.11	-10.96	1.3501	-10.82	2.6463	-11.19	-0.7381	-10.80	2.7903	-11.02	0.7741
Record 32	0.00	0.24	-0.2400	0.15	-0.1493	-0.23	0.2347	0.32	-0.3200	-0.10	0.0960

M1000 Modules H, I, J, K, L, and M Calibration Check

	module M	M %error
Record 0	-0.15	0.1493
Record 1	-0.15	0.1493
Record 2	11.06	0.4860
Record 3	110.82	0.2598
Record 4	222.13	0.0414
Record 5	333.26	0.0214
Record 6	444.19	0.0558
Record 7	555.36	0.0342
Record 8	666.69	-0.0042
Record 9	778.08	-0.0399
Record 10	888.66	0.0246
Record 11	999.73	0.0257
Record 12	1110.85	0.0236
Record 13	2221.92	0.0135
Record 14	3333.47	-0.0043
Record 15	4444.96	-0.0117
Record 16	4955.97	-0.0084
Record 17	-4956.33	-0.0156
Record 18	-4445.17	-0.0164
Record 19	-3333.89	-0.0167
Record 20	-2222.79	-0.0256
Record 21	-1111.44	-0.0297
Record 22	-1000.24	-0.0250
Record 23	-889.13	-0.0282
Record 24	-778.03	-0.0337
Record 25	-667.02	-0.0546
Record 26	-555.98	-0.0767
Record 27	-444.90	-0.1026
Record 28	-333.70	-0.1114
Record 29	-222.56	-0.1530
Record 30	-111.20	-0.0810
Record 31	-11.19	-0.7141
Record 32	0.00	0.0000