

CB7017-01 Check

	Ch0	ch0 %error	ch1	ch1 %error	Ch2	ch2 %error	Ch3	ch3 %error	Ch4	ch4 %error
Record 0	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000
Record 1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000
Record 2	0.0100	0.0000	0.0100	0.0000	0.0100	0.0000	0.0100	0.0000	0.0100	0.0000
Record 3	-0.0100	0.0000	-0.0100	0.0000	-0.0100	0.0000	-0.0100	0.0000	-0.0100	0.0000
Record 4	-1.0002	0.0039	-1.0002	0.0038	-1.0002	0.0037	-1.0002	0.0036	-1.0002	0.0034
Record 5	1.0002	-0.0032	1.0002	-0.0033	1.0002	-0.0032	1.0002	-0.0033	1.0002	-0.0032
Record 6	5.0000	0.0000	5.0000	0.0000	5.0000	0.0000	5.0000	0.0000	5.0000	0.0000
Record 7	-5.0000	0.0000	-5.0000	0.0000	-5.0000	0.0000	-5.0000	0.0000	-5.0000	0.0000
Record 8	-4.2756	0.0124	-4.2756	0.0123	-4.2756	0.0123	-4.2756	0.0126	-4.2756	0.0124
Record 9	4.2757	-0.0132	4.2757	-0.0133	4.2757	-0.0134	4.2757	-0.0135	4.2756	-0.0130

CB7017-01 Check

	Ch5	ch5 %error	Ch6	ch6 %error	Ch7	ch7 %error	Set Voltage
Record 0	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.000000
Record 1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.000000
Record 2	0.0100	0.0000	0.0100	0.0000	0.0100	0.0000	0.010000
Record 3	-0.0100	0.0000	-0.0100	0.0000	-0.0100	0.0000	-0.010000
Record 4	-1.0002	0.0035	-1.0002	0.0038	-1.0002	0.0036	-1.000000
Record 5	1.0002	-0.0034	1.0002	-0.0031	1.0002	-0.0033	1.000000
Record 6	5.0000	0.0000	5.0000	0.0000	5.0000	0.0000	5.000000
Record 7	-5.0000	0.0000	-5.0000	0.0000	-5.0000	0.0000	-5.000000
Record 8	-4.2756	0.0123	-4.2756	0.0124	-4.2756	0.0125	-4.275000
Record 9	4.2757	-0.0130	4.2757	-0.0131	4.2757	-0.0133	4.275000

CB70717-01 Check



