

M1000 Modules B C Calibration Check

	input mV	module B	B %error	module C	C %error
Record 0	0.00	0.32		0.07	
Record 1	11.11	11.39	-2.54	11.12	-0.09
Record 2	111.11	111.12	-0.01	111.13	-0.02
Record 3	222.22	222.48	-0.12	221.89	0.15
Record 4	333.33	333.57	-0.07	333.43	-0.03
Record 5	444.44	444.64	-0.05	444.03	0.09
Record 6	555.55	555.72	-0.03	555.45	0.02
Record 7	666.66	666.70	-0.01	666.70	-0.01
Record 8	777.77	778.04	-0.03	777.81	-0.01
Record 9	888.88	888.96	-0.01	888.94	-0.01
Record 10	999.99	1000.08	-0.01	999.80	0.02
Record 11	1111.11	1111.20	-0.01	1111.13	0.00
Record 12	2222.22	2221.90	0.01	2221.12	0.05
Record 13	3333.33	3332.83	0.01	3332.12	0.04
Record 14	4444.44	4444.26	0.00	4442.56	0.04
Record 15	4955.55	4955.09	0.01	4953.38	0.04
Record 16	-4955.55	-4954.66	0.02	-4954.19	0.03
Record 17	-4444.44	-4443.92	0.01	-4443.23	0.03
Record 18	-3333.33	-3332.99	0.01	-3332.35	0.03
Record 19	-2222.22	-2221.98	0.01	-2221.55	0.03
Record 20	-1111.11	-1111.09	0.00	-1111.03	0.01
Record 21	-999.99	-999.70	0.03	-999.72	0.03
Record 22	-888.88	-888.40	0.05	-888.70	0.02
Record 23	-777.77	-777.61	0.02	-777.52	0.03
Record 24	-666.66	-666.16	0.08	-666.51	0.02
Record 25	-555.55	-555.45	0.02	-555.23	0.06
Record 26	-444.44	-444.22	0.05	-444.19	0.06
Record 27	-333.33	-333.34	0.00	-333.04	0.09
Record 28	-222.22	-222.13	0.04	-222.10	0.05
Record 29	-111.11	-110.83	0.26	-110.89	0.20
Record 30	-11.11	-10.92	1.71	-10.85	2.38
Record 31	0.00	0.40		0.02	

M1000 Module B & C Calibration Check

