

## M1000 Modules g i Calibration Check

	input mV	module g	g %error	module I	i %error
Record 0	0.00	-0.04		0.10	
Record 1	11.11	10.16	8.55	10.20	8.19
Record 2	111.11	111.07	0.04	111.09	0.02
Record 3	222.22	222.17	0.02	222.26	-0.02
Record 4	333.33	333.23	0.03	333.41	-0.02
Record 5	444.44	444.34	0.02	444.44	0.00
Record 6	555.55	555.40	0.03	555.51	0.01
Record 7	666.66	666.65	0.00	666.54	0.02
Record 8	777.77	777.55	0.03	777.75	0.00
Record 9	888.88	888.72	0.02	888.81	0.01
Record 10	999.99	999.59	0.04	999.84	0.02

### M1000 Modules g & i Calibration Check

