

Temperature Calibration Report

Customer:	Oregon State University			
Job Number:	79487	Date of Report:	7/17/2014	
Model Number	SBE 03-01/F	Serial Number:	030997	

Temperature sensors are normally calibrated 'as received', without adjustments, allowing a determination sensor drift. If the calibration identifies a problem, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request.

An 'as received' calibration certificate is provided, listing coefficients to convert sensor frequency to temperature. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients. The coefficient 'offset' allows a small correction for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair apply only to subsequent data.

'AS RECEIVED CALIBRATION'		Performed	□ Not Performed
Date: 7/2/2014	Drift since last cal	-0.00302	Degrees Celsius/year
Comments:			
'FINAL CALIBRATION'		Performed	Not Performed
Date: 7/16/2014	Drift since 31 Jul	-0.0032	3 Degrees Celsius/year
Comments:			
The connector was upgraded to wet-pluggable s	style.		