

SEA-BIRD ELECTRONICS, INC.

1900 NE 20th Street, Bellevue, Washington, 98005-2011

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1320

CALIBRATION DATE: 30-Jun-11

ITS-90 COEFFICIENTS

$g = 4.82727135e-003$

$h = 6.77223932e-004$

$i = 2.61453523e-005$

$j = 2.01650977e-006$

$f_0 = 1000.0$

BATH TEMP
(ITS-90)

INSTRUMENT FREQ
(Hz)

-1.4998	6049.790
1.0002	6396.813
4.5002	6906.409
8.0002	7444.426
11.5002	8011.618
15.0002	8608.695
18.5002	9236.381
22.0002	9895.338
25.5002	10586.216
29.0002	11309.669
32.5002	12066.279

Temperature ITS-90 = $1 / \{ g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)] \}$

Temperature IPTS-68 = $1 / \{ a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)] \}$

Following the recommendation of JPOTS: T_{68} is assumed

Residual = instrument temperature - bath temperature

Date, Offset(mdeg C)

