Temperature Calibration Report

Customer:	Oregon State Ur	niversity			
Job Number:	79487	Da	te of Report	::	7/17/2014
Model Number	SBE 03-01/F	Sei	rial Number	:	031366
Temperature sensors are normally calibrated 'as received', without adjustments, allowing a determination sensor drift. If the calibration identifies a problem, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request. An 'as received' calibration certificate is provided, listing coefficients to convert sensor frequency to temperature. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients. The coefficient 'offset' allows a small correction for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair apply only to subsequent data. 'AS RECEIVED CALIBRATION' Performed Not Performed					
Date: 7/2/2014		Drift since	ast cal:	+0.00202	Degrees Celsius/year
Comments:					
'FINAL CALIBRA	ATION'		✓ Perfo	rmed	□ Not Performed
Date: 7/16/2014]	Drift since	18 Jul 13	+0.0002	Degrees Celsius/year

Comments:

The connector was upgraded to wet-pluggable style.