Comments:

## **Temperature Calibration Report**

Customer:	Oregon State Unive	rsity		
Job Number:	65123	Date of Repo	ort:	7/19/2011
Model Number	SBE 03-02/F	Serial Numb	er:	031384
Temperature sensors are normally calibrated 'as received', without adjustments, allowing a determination sensor drift. If the calibration identifies a problem, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request.  An 'as received' calibration certificate is provided, listing coefficients to convert sensor frequency to temperature. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients using the program SEACON. The coefficient 'offset' allows a small correction for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair apply only to subsequent data.				
'AS RECEIVED	CALIBRATION'	✓ Per	formed	☐ Not Performed
Date: 7/19/201	1	Drift since last cal:	-0.0006	Degrees Celsius/year
Comments:				
'CALIBRATION	AFTER REPAIR'	□ Per	formed	✓ Not Performed
Date:		Drift since Last cal:		Degrees Celsius/year