Temperature Calibration Report

Customer:	Oregon State Ur	niversity			
Job Number:	79487		Date of Rep	ort:	7/17/2014
Model Number	SBE 03-02/F		Serial Numb	er:	031384
Temperature sensors are normally calibrated 'as received', without adjustments, allowing a determination sensor drift. If the calibration identifies a problem, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request. An 'as received' calibration certificate is provided, listing coefficients to convert sensor frequency to temperature. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients. The coefficient 'offset' allows a small correction for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair apply only to subsequent data.					
'AS RECEIVED CALIBRATION'			✓ Pei	rformed	☐ Not Performed
Date: 7/2/2014]	Drift sinc	e last cal:	-0.00009	Degrees Celsius/year
Comments:					
'FINAL CALIBRA	ATION'		✓ Per	rformed	□ Not Performed
Date: 7/16/2014]	Drift sinc	e 31 Jul 12	+0.000	Degrees Celsius/year

Comments:

The connector was upgraded to wet-pluggable style.