

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1384
CALIBRATION DATE: 16-Jul-14

SBE 3 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS:

g = 4.86995503e-003
h = 6.80569185e-004
i = 2.70883961e-005
j = 2.15085794e-006
f0 = 1000.0

BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5000	6453.288	-1.5000	-0.00001
1.0000	6823.907	1.0000	0.00002
4.5000	7368.167	4.5000	0.00002
8.0000	7942.787	8.0000	-0.00004
11.5000	8548.586	11.5000	-0.00000
15.0000	9186.310	15.0000	0.00001
18.5000	9856.708	18.5000	0.00002
22.0000	10560.490	22.0000	0.00002
25.5000	11298.337	25.5000	-0.00002
29.0000	12070.922	29.0000	-0.00004
32.5000	12878.888	32.5000	0.00002

Temperature ITS-90 = $1/\{g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]\} - 273.15$ (°C)

Residual = instrument temperature - bath temperature

