



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Oregon State University

Job Number 79487

Date 7/2/2014

Technician BLT

Serial Number 031384

Low Pressure (PSI) 50 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) 5000 PSI

Time (Minutes) 30 Minutes

Pass

Fail

Comments

Replaced the main piston "O"-Rings. Replaced 3-pin connector with 3-pin wet-pluggable.

