

# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2327  
CALIBRATION DATE: 21-Jul-10

SBE3 TEMPERATURE CALIBRATION DATA  
ITS-90 TEMPERATURE SCALE

## ITS-90 COEFFICIENTS

g = 4.35183855e-003  
h = 6.42766446e-004  
i = 2.32333095e-005  
j = 2.24439028e-006  
f0 = 1000.0

## IPTS-68 COEFFICIENTS

a = 3.68121358e-003  
b = 6.00543780e-004  
c = 1.59808709e-005  
d = 2.24592846e-006  
f0 = 2948.184

BATH TEMP (ITS-90)	INSTRUMENT FREQ (Hz)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5001	2948.184	-1.5001	0.00001
0.9999	3117.974	0.9999	-0.00001
4.4999	3367.420	4.4999	0.00000
7.9999	3630.908	7.9999	-0.00001
11.4999	3908.820	11.4999	0.00000
14.9999	4201.525	15.0000	0.00007
18.5000	4509.366	18.4999	-0.00006
22.0000	4832.702	22.0000	-0.00001
25.5000	5171.854	25.5000	0.00002
29.0000	5527.138	29.0000	0.00001
32.5000	5898.861	32.5000	-0.00001

Temperature ITS-90 =  $1/\{g + h[\ln(f_0/f)] + i[\ln^2(f_0/f)] + j[\ln^3(f_0/f)]\} - 273.15$  (°C)

Temperature IPTS-68 =  $1/\{a + b[\ln(f_0/f)] + c[\ln^2(f_0/f)] + d[\ln^3(f_0/f)]\} - 273.15$  (°C)

Following the recommendation of JPOTS:  $T_{68}$  is assumed to be  $1.00024 * T_{90}$  (-2 to 35 °C)

Residual = instrument temperature - bath temperature

