

Temperature Calibration Report

Customer:	Oregon State Ur	niversity	
Job Number:	51842	Date of Report:	9/4/2008
Model Number	SBE 03S	Serial Number:	032767

Temperature sensors are normally calibrated 'as received', without adjustments, allowing a determination sensor drift. If the calibration identifies a problem, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request.

An 'as received' calibration certificate is provided, listing coefficients to convert sensor frequency to temperature. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients using the program SEACON. The coefficient 'offset' allows a small correction for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair apply only to subsequent data.

'AS RECEIVED CALIBRATION'	\checkmark Performed \Box Not Perform	ned
Date: 9/4/2008	Drift since last cal: -0.00036 Degrees Celsius	s/year
Comments:		
'CALIBRATION AFTER REPAIR'	□ Performed ✓ Not Perform	ned
Date:	Drift since Last cal: Degrees Celsius	
Comments:		