



# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

## Pressure Test Certificate

Customer Oregon State University

Job Number 63973

Date 5/25/2011

Technician BLT

Serial Number 3215736-0194

Low Pressure (PSI) 50 PSI

Time (Minutes) 30 Minutes

High Pressure (PSI) N/A\* PSI

Time (Minutes) N/A\* Minutes

Pass

Fail

Comments

Replaced the main piston "O"-Rings.  
\* Due to the dimensions of the CAROUSEL we are unable to perform a high pressure test at this time.

