



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Oregon State University

Job Number 77172

Date 3/11/2014

Technician BLT

Serial Number 3215736-0194

Low Pressure (PSI) 50 PSI

Time (Minutes) 60 Minutes

High Pressure (PSI) * N/A PSI

Time (Minutes) * N/A Minutes

Pass

Fail

Comments

Replaced the main piston "O"-Rings. Replaced 6-pin connector with 6-pin wet-pluggable connector.

*Due to the dimensions of the CAROUSEL we are unable to perform a high pressure test at this time.

