

# SBE38

---

*Digital Oceanographic Thermometer*

## Instrument Configuration:

Serial Number	3856968-0607
Firmware Version	1.4
Interface Type	RS-232
Baud Rate	4800
<b>Maximum Depth</b>	<b>10500 meters</b>

**CAUTION - The maximum deployment depth is limited by the pressure housing to 10500m.**



Sea-Bird Electronics, Inc.

13431 NE 20th St. Bellevue, Washington 98005 USA

Website: <http://www.seabird.com>

Phone: (425) 643-9866

FAX: (425) 643-9954

Email: [seabird@seabird.com](mailto:seabird@seabird.com)

## SBE Pressure Test Certificate

Test Date: 8/12/2010 Description SBE-38 Digital Thermometer

Job Number: 56968 Customer Name OSU

### SBE Sensor Information:

Model Number: 38

Serial Number: 0607

### Pressure Sensor Information:

Sensor Type: None

Sensor Serial Number: None

Sensor Rating: 0

### Pressure Test Protocol:

Low Pressure Test: 40 PSI Held For: 15 Minutes

High Pressure Test: 10000 PSI Held For: 15 Minutes

Passed Test:

Tested By: VG

