



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Oregon State University

Job Number 68054

Date 5/11/2012

Technician TK

Serial Number 3856968-0607

Low Pressure (PSI) 45 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) 10000 PSI

Time (Minutes) 15 Minutes

Pass

Fail

Comments

Replaced temperature probe. Replaced the main piston "O"-Rings.

