

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0607
CALIBRATION DATE: 21-Jan-15

SBE 38 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 6.757767e-005
a1 = 2.699130e-004
a2 = -2.142661e-006
a3 = 1.435324e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
-1.5000	747610.9	-1.5000	0.0000
1.0000	665970.0	1.0000	0.0000
4.5000	568104.3	4.5000	-0.0000
8.0000	486248.4	8.0000	-0.0000
11.5000	417543.6	11.5000	0.0000
15.0000	359682.9	15.0000	0.0000
18.5000	310793.6	18.5000	0.0000
22.0000	269352.8	22.0000	0.0000
25.5000	234116.1	25.5000	-0.0000
29.0000	204063.5	28.9999	-0.0001
32.5000	178355.9	32.5001	0.0001

Temperature ITS-90 = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$ (°C)

Residual = instrument temperature - bath temperature

n = instrument output

