

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA
 Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0497
 CALIBRATION DATE: 25-Feb-14

SBE4 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -4.25944329e+000
 h = 4.59733608e-001
 i = -2.75279747e-004
 j = 3.23198936e-005
 CPcor = -9.5700e-008 (nominal)
 CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 2.71540964e-006
 b = 4.58812661e-001
 c = -4.25641292e+000
 d = -8.72791888e-005
 m = 4.7
 CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	3.04564	0.00000	0.00000
-1.0000	34.8083	2.80399	8.38226	2.80400	0.00001
1.0000	34.8087	2.97538	8.60121	2.97537	-0.00001
15.0000	34.8083	4.27074	10.10202	4.27074	0.00001
18.5000	34.8080	4.61739	10.46662	4.61739	-0.00000
29.0000	34.8065	5.70089	11.53066	5.70089	-0.00001
32.5000	34.7999	6.07345	11.87401	6.07346	0.00001

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$\text{Conductivity} = (af^m + bf^2 + c + dt) / [10(1 + \epsilon p)] \text{ Siemens/meter}$$

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

