

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0670
CALIBRATION DATE: 17-Nov-06

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -4.30979201e+000
h = 4.57263175e-001
i = 4.17904329e-004
j = 6.38255946e-007
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 5.94345727e-004
b = 4.56770647e-001
c = -4.30826910e+000
d = -1.04498169e-004
m = 2.9
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	3.06573	0.00000	0.00000
-0.9971	34.9671	2.81583	8.39392	2.81584	0.00001
14.9999	34.9675	4.28818	10.11179	4.28812	-0.00006
18.4999	34.9652	4.63597	10.47617	4.63603	0.00006
28.9999	34.9619	5.72346	11.54016	5.72346	-0.00001
32.4999	34.9577	6.09784	11.88413	6.09784	-0.00000

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10 (1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

