

# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0670  
CALIBRATION DATE: 09-Feb-11

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## GHIJ COEFFICIENTS

g = -4.30892623e+000  
h = 4.56822412e-001  
i = 5.14983571e-004  
j = -4.50517829e-006  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 5.91837703e-004  
b = 4.56817310e-001  
c = -4.30968933e+000  
d = -8.04187587e-005  
m = 2.9  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	3.06607	0.00000	0.00000
-1.0000	34.9765	2.81627	8.39425	2.81627	-0.00000
1.0000	34.9766	2.98836	8.61294	2.98836	0.00000
15.0000	34.9755	4.28907	10.11254	4.28907	-0.00000
18.5000	34.9741	4.63704	10.47698	4.63704	0.00000
29.0001	34.9665	5.72415	11.54080	5.72415	0.00000

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

