

# SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0830  
CALIBRATION DATE: 17-Nov-06

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## GHIJ COEFFICIENTS

g = -4.11034862e+000  
h = 5.29063357e-001  
i = -4.96034382e-004  
j = 5.45691483e-005  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 2.56579047e-006  
b = 5.27465930e-001  
c = -4.10578240e+000  
d = -1.19823701e-004  
m = 4.9  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.78984	0.00000	0.00000
-0.9971	34.9671	2.81583	7.81377	2.81585	0.00002
14.9999	34.9675	4.28818	9.42310	4.28809	-0.00009
18.4999	34.9652	4.63597	9.76407	4.63604	0.00007
28.9999	34.9619	5.72346	10.75857	5.72351	0.00005
32.4999	34.9577	6.09784	11.07965	6.09780	-0.00004

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

