

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1018
CALIBRATION DATE: 03-Jul-14

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -4.08991744e+000
h = 4.62869866e-001
i = -4.57498277e-004
j = 4.74843870e-005

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.97556	0.00000	0.00000
-1.0000	34.7200	2.79754	8.32784	2.79757	0.00003
1.0000	34.7203	2.96854	8.54628	2.96852	-0.00002
15.0000	34.7213	4.26119	10.04273	4.26110	-0.00009
18.5000	34.7207	4.60705	10.40610	4.60713	0.00008
29.0000	34.7197	5.68827	11.46545	5.68831	0.00003
32.5000	34.7116	6.05979	11.80679	6.05977	-0.00003

f = INST FREQ / 1000.0

Conductivity = (g + h * f² + i * f³ + j * f⁴) / (1 + δ * t + ε * p) Siemens / meter

t = temperatur e[°C]; p = pressure[decibars]; δ = CTcor; ε = CPcor;

Residual = instrument conductivity - bath conductivity

