

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1018
CALIBRATION DATE: 09-Jul-14

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -4.09051163e+000
h = 4.63014487e-001
i = -4.86173648e-004
j = 4.89346040e-005

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.97555	0.00000	0.00000
-1.0000	34.7531	2.79996	8.33099	2.79999	0.00003
0.9999	34.7536	2.97111	8.54955	2.97109	-0.00001
15.0000	34.7542	4.26480	10.04661	4.26470	-0.00010
18.5000	34.7537	4.61096	10.41015	4.61103	0.00007
29.0000	34.7541	5.69328	11.47012	5.69333	0.00006
32.5000	34.7494	6.06564	11.81202	6.06560	-0.00005

f = INST FREQ / 1000.0

Conductivity = (g + h * f² + i * f³ + j * f⁴) / (1 + δ * t + ε * p) Siemens / meter

t = temperatur e[°C]; p = pressure[decibars]; δ = CTcor; ε = CPcor;

Residual = instrument conductivity - bath conductivity

