

# Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1029  
CALIBRATION DATE: 25-Feb-14

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## GHIJ COEFFICIENTS

g = -3.98420429e+000  
h = 5.40076620e-001  
i = -6.76534038e-005  
j = 3.31072500e-005  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 2.20794475e-005  
b = 5.39909874e-001  
c = -3.98391408e+000  
d = -8.63432693e-005  
m = 4.1  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.71593	0.00000	0.00000
-1.0000	34.8083	2.80399	7.69013	2.80400	0.00000
1.0000	34.8087	2.97538	7.89257	2.97537	-0.00000
15.0000	34.8083	4.27074	9.27917	4.27073	-0.00000
18.5000	34.8080	4.61739	9.61585	4.61739	-0.00000
29.0000	34.8065	5.70089	10.59812	5.70090	0.00001
32.5000	34.7999	6.07345	10.91499	6.07345	-0.00001


Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



25-Feb-14 1.0000000