

# SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1030  
CALIBRATION DATE: 17-Nov-06

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## GHIJ COEFFICIENTS

g = -4.06778998e+000  
h = 5.73057721e-001  
i = 2.83644917e-005  
j = 3.03441576e-005  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 4.09446038e-005  
b = 5.73111270e-001  
c = -4.06802818e+000  
d = -1.12838108e-004  
m = 3.9  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.66360	0.00000	0.00000
-0.9971	34.9671	2.81583	7.48654	2.81584	0.00001
14.9999	34.9675	4.28818	9.03007	4.28812	-0.00006
18.4999	34.9652	4.63597	9.35717	4.63603	0.00006
28.9999	34.9619	5.72346	10.31168	5.72346	0.00000
32.4999	34.9577	6.09784	10.62005	6.09783	-0.00001

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

